

6640 185th Ave NE, Redmond, WA 98052

T.425.895.8617, F.425.702.9358

ON-SITE CERTIFICATE #: OS3-00335468

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

KITSAP COUNTY SHERIFFS OFFICE

3951 NW RANDALL WAY SILVERDALE, WA 98383

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2017, that is As Found data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRIPTION:		SPEED MEASURING DEVICE, DOPPLER					
ASSET #:	142		SERIAL NUMBER:		RP02448		
MANUFACTURER:	KUSTOM SIGNALS, INC.		MODEL NUMBER:		RAPTOR RP-1		
DEPARTMENT:	N/A		PATROL CAR #:		N/A		
ENVIRONMENT:	70.6 °F/45.5 %RH		BASIC ACCURACY:	REFERENCE MFRS SPECIFICATION			
CAL INTERVAL:	24 MONTHS		DUE DATE:	-	January 03, 2027		
EQUIPMENT CONDITION AS RECEIVED Initial testing found this equipment to be "IN TOLERANCE", as defined by the basic accuracy stated above. EQUIPMENT CONDITION AS DELIVERED							
At the completion of the calibration, measured values were "IN TOLERANCE", as defined by the basic accuracy stated above.							
TUNNING FORK(S) SUPPLIED WITH THIS DEVICE							
DESCRIPT		SERIAL NUMBER	RATED SPEED		FREQUE	NCY	
TUNING FORK ONE TUNING FORK TWO	4739 4463			35 M.P.H. 65 M.P.H.		2548 4744	
Antenna 1 SN: RK03843, Frequency: 24.214 GHz Antenna 2 SN: RK03838, Frequency: 24.191 GHz							
STANDARD(S) USED FOR CERTIFICATION							
	MODEL	MANUFACTURER	DESC	DESCRIPTION		DUE DAT	
MET1231 VOCAR H		DB INNOVATIONS	HAND HELD RADAR CERTIFICATION SYSTEM		TEM	03/26/2025	
MET1232 VOCAR H	R WAND	DB INNOVATIONS	VOCAR HR WAND			03/26/2025	
PROCEDURE(S) USED FOR CERTIFICATION							
DOCUMENT ID			ESCRIPTION		REV	- Andreas - Andr	
CP-SMD-001	RADAR AND L	IDAR CALIBRATION PROCED	URE		NEW	11/14/2019	
CERTIFICATION	INOȚES						
I certify (or declare) under penalty of perjury under the laws of the State of Washington that the above information is true and correct							
LOCATION: SILVERDA						.E, WA	
PERFORMED BY_	SEMIO	R METROLOGIST: JOHN GRAY	CALIBRATION	N DATE:	Friday, January	03, 2025	
THIS LABORATORY IS A2LA	ACCREDITED TO ISO/IEC	17025:2017 (GENERAL REQUIREMENT	S FOR THE COMPETENCE OF TESTING AN	ID CALIBRATION	LABORATORIES), CERTI	FICATE #: 2560.01	